

## Perbedaan detil gambaran radiografis obyek interdental pada model gigi menggunakan film instan dengan dan tanpa filter

Rio Suryantoro, author

Deskripsi Lengkap: <https://lib.ui.ac.id/detail?id=128124&lokasi=lokal>

---

### Abstrak

Nowadays, instant film is being an alternative choice on supportive radiographic examination in dentistry. Early diagnosis of alteration in periodontal tissue, especially gingiva which has a low density, on a conventional radiographic image often give a difficulty to the dentists. Detail image of soft tissue can be gained by decreasing radiographic contrast. Added aluminium filter itself can influence radiographic contrast. The purpose of this research is to observe the effect of coin Rp. 100,- as added aluminium filter in increasing the detail of intraoral image of interdental object with D-speed instant film Hanshin© on phantom. Sixty radiographic images were taken. 30 films added with aluminium filter and 30 films without any added filter. The examination was taken by measuring the height of vertical dimension of interdental papilla's opacity with caliper. Data gained in the region of interest which was the interdental papilla between mandibular right first molar and mandibular second molar were analyze with independent-sample t test. The result showed that there is differences in mean of the height of interdental object's opacity ( $p < 0,05$ ). It was concluded that coin Rp. 100,- as added aluminium filter can increase the radiographic detail of interdental object with instant film compare with the intraoral image without any added filtration.

Dewasa ini, penggunaan film instan menjadi pilihan alternatif dalam pemeriksaan penunjang radiografis di bidang kedokteran gigi. Deteksi dini perubahan jaringan periodonsium pada radiograf konvensional sering memberikan kesulitan bagi dokter gigi, terutama jaringan lunak/gingiva yang mempunyai densitas minimal. Gambaran detil jaringan lunak dapat diraih dengan cara menurunkan kekontrasan radiograf. Penggunaan filter aluminium tambahan dapat mempengaruhi kekontrasan radiograf. Penelitian ini bertujuan untuk melihat pengaruh uang logam seratus rupiah sebagai filter dalam meningkatkan detil gambaran radiografis obyek interdental model gigi pada foto intraoral dengan film instan Hanshin D-speed. Obyek penelitian adalah 60 buah radiograf, 30 tanpa menggunakan filter aluminium dan 30 radiograf menggunakan filter aluminium. Pemeriksaan detil dilakukan dengan cara mengukur panjang dimensi vertikal opasitas papila interdental menggunakan kaliper (milimeter). Data yang diperoleh adalah data numerik. Regio yang diperiksa adalah obyek interdental antara gigi molar 1 bawah kanan dengan gigi molar 2 bawah kanan. Hasil penelitian ini menunjukkan bahwa terdapat perbedaan rerata panjang dimensi vertikal opasitas obyek interdental (uji t tidak berpasangan,  $p < 0,05$ ). Dari penelitian ini dapat disimpulkan bahwa filter aluminium berupa uang logam Rp 100,- (seratus rupiah) terbukti dapat meningkatkan detil obyek interdental model gigi foto intraoral dengan film instan dibandingkan dengan foto tanpa menggunakan filter.