

Hubungan antara quick of blood (Qb) dengan adekuasi hemodialisis pasien yang menjalani terapi remodialisis di ruang HD BRSU faerah Tamanan Bali

I Gusti Ayu Puja Astuti Dewi, author

Deskripsi Lengkap: <https://lib.ui.ac.id/detail?id=137264&lokasi=lokal>

Abstrak

ABSTRAK

Selama proses hemodialisis darah dari tubuh pasien dialirkan menuju sirkuit darah . Pengaturan kecepatan aliran darah diatur oleh pompa darah sebagai Quick of Blood/Qb. Penelitian ini bertujuan untuk mengetahui hubungan antara Qb dengan adekuasi hemodialisis pada pasien yang menjalani terapi hemodialisis.

Desain penelitian menggunakan cross-sectional dengan 38 orang responden. Qb diobservasi dengan melihat nilai yang tercantum pada mesin dan adekuasi hemodialisis dihitung dengan menggunakan rumus turunan pertama Kt/V dan URR. Hubungan antara Qb dengan adekuasi hemodialisis diuji dengan korelasi dan regresi linier. Rata-rata nilai Qb adalah 222,94 mL/menit dengan standar deviasi/SD 23,17. Rata-rata nilai adekuasi hemodialisis yaitu 1,22 (62,18%) dengan SD 0,34.

Hasil uji statistik menunjukkan tidak ada hubungan bermakna antara Qb dengan adekuasi hemodialisis (p value = 0,225). Kepatenaan akses vaskuler, berat badan, komplikasi intradialisis, dan ukuran lumen kateter dapat digunakan sebagai pedoman dalam pengaturan Qb. Perawat perlu memperhatikan pedoman pengaturan Qb ini untuk mencapai adekuasi hemodialisis yang optimal.

<hr><i>ABSTRACT

During haemodialysis, blood is transferred from patient into a blood circuit. The adjustment of blood stream is regulated by a blood pump as quick blood/Qb. The aim of this study was to identify the relationship between Qb and haemodialysis adequacy among patient undergoing haemodialysis.

Cross sectional design with 38 respondents were adopted in this study. Qb was evaluated by observing the indicator at the machine and haemodialysis adequacy was calculated with the first derivative formula of Kt/V and URR. The relationship between Qb and haemodialysis adequacy was tested by correlation analyses and linier regression. Mean Qb was 222.94 mL/minute with standard deviation of 23.17. Adequacy value was 1.22 (62,18%) with standard deviation of 0.34.

Statistical analyses showed no significant relationships between Qb and haemodialysis adequacy (p value = 0.225). Adequacy vascular access, weight, intra-dialysis complications, and lumen catheter size could be applied as a guideline in the adjustment of Qb. Therefore, nurses need to pay attention on this guideline to reach the optimum adequacy.</i>