

# Estimasi dosis radiasi pada pemeriksaan CT scan kepala = Radiation dose estimation in head CT scan examination

Elia Soediatmoko, author

Deskripsi Lengkap: <https://lib.ui.ac.id/detail?id=20290765&lokasi=lokal>

---

## Abstrak

Pemeriksaan CT scan kepala sudah menjadi pemeriksaan rutin untuk kasus sakit kepala. Namun informasi dosis radiasi pemeriksaan CT scan kepala belum banyak diketahui. Informasi akan dosis ini sangatlah penting karena adanya organ yang sensitif terhadap radiasi seperti kelenjar thyroid, kelenjar air ludah, lensa mata dan otak kepala. Untuk mengetahui estimasi nilai dosis di organ kepala tersebut digunakan software ImPACT CT patient Dosimetry Calculator yang menggunakan nilai nCTDIw yang diperoleh dari hasil pengukuran menggunakan detektor bilik ionisasi pensil berukuran 100 mm dengan obyek phantom CTDI berukuran 160 mm sebagai salah satu faktor penghitungan. Dari 15 pasien diestimasi dosis ekivalen untuk dosis ekivalen thyroid 0.072 mSv - 0.33 mSv, Kelenjar air ludah berkisar 0.66 mSv - 0.8 mSv, otak kepala 0.66 mSv - 0.8 mSv, Sedang untuk lensa mata dinyatakan dalam dosis organ karena alasan deterministik kemungkinan terjadinya katarak pada lensa mata karena radiasi, yakni sebesar 75 mGy - 91 mGy, serta total dosis efektif 3 mSv - 3.7 mSv, pada parameter uji 120 kV 300 mAs. Besar nilai dosis dipengaruhi oleh mAs, panjang scan dan pitch, sehingga proteksi radiasi terhadap organ thyroid harus dilakukan.

.....Head CT scan has become a rutin procedure to rule out headache symptoms, but dose radiation influences is yet to be known . Information dose of head CT scan is very important because there are organ at risk such thyroid, saliva glands,brain and eye lens. Using nCTDIw values obtained from the measurement of 100 mm pencil ionization chamber on 16 cm CTDI phantom, combined with 15 patient data obtained from DICOM data patient, and estimated dose using imPACT CT patient dose calculator, estimated equivalent dose are, for thyroid 0.072 mSv - 0.33 mSv, saliva glands 0.66 mSv - 0.8 mSv, brain 0.66 mSv- 0.8 mSv and the eye lens are mention in organ dose because of deterministic reason of cataract formation rather than for effective dose calculation are 75 mGy - 91 mGy and estimated total dose effective are 3 mSv - 3.71 mSv at 120 kV 300mAs. The dose value is influenced by mAs, lenght of scan and pitch, for futher attention of radiation protection for thyroid gland area must be done.