

Perbandingan efektivitas dan keamanan fakoemulsifikasi torsional dan transversal terhadap densitas sel endotel kornea pada katarak sedang keras = Comparative of effectivity and safety between torsional and transversal phacoemulsification againts corneal endothelial cell density in moderate hard cataract / Nashrul Ihsan

Nashrul Ihsan, author

Deskripsi Lengkap: <https://lib.ui.ac.id/detail?id=20423415&lokasi=lokal>

Abstrak

ABSTRAK

Tujuan Untuk mengetahui perbandingan efektivitas dan keamanan antara fakoemulsifikasi torsional dan transversal menggunakan parameter phaco time dan perubahan sel endotel dan ketebalan kornea sentral Metode Penelitian prospektif menggunakan katarak senilis NO 3 4 LOCS III yang dilakukan randomisasi menjadi dua kelompok torsional Ozil IP dan transversal Ellips FX Keluaran primer berupa phaco time sel endotel kornea ketebalan kornea sentral tajam penglihatan terkoreksi pada hari pertama ketujuh dan ke 30 pasca operasi Hasil Penelitian ini menggunakan 61 pasient Karakteristik dasar setara dan dapat dibandingkan Phaco time torsional CDE memiliki nilai kecil hingga hanya 1 3 phaco time transversal EFX Penurunan ECD kelompok torsional 7 9 dan kelompok transversal 8 9 Tidak ada perbedaan bermakna pada perubahan ECD dan CCT antara fakoemulsifikasi torsional dan transversal Simpulan Efektivitas dan keamanan kedua mesin fakoemulsifikasi torsional dan transversal tidak berbeda signifikan.

<hr><i>ABSTRACT</i>

Purpose To compare the effectivity and safety between torsional and transversal phacoemulsification using intraoperative parameter phaco time and postoperative parameter endothelial cells and central corneal thickness changesMethods This prospective study with senile cataract eyes NO 3 4 LOCS III which randomized to have phacoemulsification using torsional Ozil IP or transversal Ellips FX Primary outcomes were phaco time endothelial cell density ECD central corneal thickness CCT corrected distance visual acuity with 1 7 and 30 days after phacoemulsificationResults The study included 61 patients Baseline characteristic were comparable The phaco time torsional CDE only one third of phaco time transversal EFX The results of the percentage of ECD loss were 7 9 in torsional and 8 9 in transversal No difference in ECD and CCT changes between torsional and transversal statistically Conclusions The effectivity and safety of torsional and transversal phacoemulsification did not differ significantly.</i>