

Kapasitas difusi paru (DLCO) pada pasien tumpang tindih asma - penyakit paru obstruktif kronik di Rumah Sakit Rujukan Respirasi Nasional Persahabatan Jakarta = Diffusing capacity of lung for carbon monoxide (DLCO) of asthma-chronic obstructive pulmonary disease overlap patients treated in Persahabatan Hospital Jakarta, Indonesia

Anna Yusrika, author

Deskripsi Lengkap: <https://lib.ui.ac.id/detail?id=20502480&lokasi=lokal>

Abstrak

ABSTRAK

Latar belakang: Kapasitas difusi paru berdasarkan karbon ke sirkulasi pulmoner. Nilai DLCO prediksi pada asma cenderung normal atau sedikit monoksida (DLCO) didesain untuk mengukur laju perpindahan gas CO dari alveolus meningkat sedangkan pada PPOK kapasitas difusi cenderung menurun akibat emfisema. Sindrom tumpang-tindih asma-PPOK dinyatakan sebagai entitas yang unik dengan kombinasi karakteristik asma dan PPOK. Tujuan utama penelitian ini adalah mengetahui nilai DLCO pada pasien tumpang tindih asma- PPOK (TAP) di RSUP Persahabatan Jakarta.

Metode: Uji DLCO dengan metode napas tunggal dan beberapa pemeriksaan penunjang lainnya telah dilakukan pada 40 pasien yang terdiagnosis sebagai TAP. Diagnosis TAP pada subjek penelitian ditegakkan menggunakan kriteria pedoman GINA/GOLD 2017. Kriteria akseptabilitas dan reproduibilitas DLCO napas tunggal dinilai menggunakan kriteria dari ATS/ERS 2017. Hasil uji DLCO disajikan dalam nilai mutlak dan nilai persen prediksi.

Hasil: Rerata nilai DLCO mutlak dan %DLCO prediksi yang didapatkan dalam penelitian ini adalah 17.98 ± 5.37 mL/menit/mmHg dan $84.16 \pm 18.29\%$. Jika menggunakan persamaan penyesuaian DLCO berdasarkan kadar hemoglobin didapatkan nilai %DLCO prediksi sedikit meningkat dibanding sebelumnya, $85.17 \pm 18.04\%$. Terdapat 10 subjek (25.0%) yang mengalami penurunan nilai DLCO. Enam diantaranya mengalami penurunan ringan dan empat lainnya mengalami penurunan sedang.

Kesimpulan : Rerata nilai DLCO pada subjek TAP di RSUP Persahabatan Jakarta dapat diinterpretasikan normal, lebih menyerupai asma dibandingkan PPOK. Hasil ini juga mengindikasikan kebanyakan pasien TAP dalam penelitian ini tidak mengalami penurunan luas permukaan alveolar yang mengganggu proses difusi.

<hr>

ABSTRACT

Background: Diffusing capacity of the lung for carbon monoxide (DLCO) was designed to measure transfer rate of carbon monoxide from alveoli to pulmonary circulation. As we know, DLCO predicted value in asthma proved to be normal or slightly elevated. On contrary it decreased in COPD with emphysematous pattern. Asthma-chronic obstructive pulmonary disease overlap (ACO) declared as a unique entity with combined characteristics between asthma and COPD. The aim of the research is to find out DLCO value of ACO patient in Persahabatan Hospital, Jakarta.

Method: We have conducted single-breath DLCO and other required test to 40 patients diagnosed with ACO using GINA/GOLD 2017 guidelines. The acceptability and reproducibility of single-breath DLCO was done according to ATS/ERS 2017 criteria. The result then presented as absolute value and percent predicted

value.

Results: The mean DLCO of our patient is 17.98 ± 5.37 mL/minute/mmHg with percent predicted value is $84.16 \pm 18.29\%$. Using adjusted DLCO equation for hemoglobin, we found that the value is slightly increased, $85.17 \pm 18.04\%$. However, we found 10 patient (25.0%) with DLCO decrease. Six of them have DLCO predicted value $<75\%$ (mild-decrease) and four of them have DLCO predicted value $<60\%$ (moderate-decrease).

Conclusion: The mean DLCO value of patient with ACO in our hospital can be interpreted as normal, similar with asthma, rather than COPD. It also indicate most of our patient did not have alveolar loss that altering diffusion process.