

Korelasi antara nilai USG Point - Shear Wave Elastography (pSWE) dan nilai USG 2 Dimension - Shear Wave Elastography (2D-SWE) pada hepatocellular carcinoma = The Correlation between the value of USG Point - Shear Wave Elastography (pSWE) and the value of 2 dimension USG Value - Shear Wave Elastography (2D-SWE) in hepatocellular carcinoma

Hafidz Muhammad Prodjokusumo, author

Deskripsi Lengkap: <https://lib.ui.ac.id/detail?id=20515344&lokasi=lokal>

Abstrak

Latar Belakang: Kanker hati adalah penyebab paling umum kedua kematian akibat kanker di seluruh dunia, dan kejadian keganasan hati 'primer' telah meningkat secara signifikan, terutama Hepatocellular Carcinoma/ HCC. Karakteristik jaringan pada keganasan biasanya lebih padat/ kaku dibandingkan jaringan normal. Ultrasonografi dengan teknologi shear wave (gelombang geser) adalah metode perhitungan kekakuan jaringan, saat ini terdapat dua tipe; elastografi gelombang geser titik (pSWE) dan elastografi gelombang geser dua dimensi (2D-SWE). Belum ada studi yang menggunakan dua tipe gelombang geser tersebut pada lesi HCC yang sama untuk melihat kesesuaian antara nilai keduanya.

Tujuan: Mengetahui derajat korelasi antara nilai USG pSWE dan nilai USG 2D-SWE pada HCC.

Metode: Sebanyak 17 subjek penelitian dengan diagnosis HCC dilakukan pemeriksaan USG pSWE dan dilanjutkan dengan 2D-SWE (pada hari yang sama atau maksimal tujuh hari setelahnya) pada lesi HCC untuk menentukan nilai kekakuan jaringan (dalam satuan kPa dan m/s). Setelah itu dilakukan analisis korelasi antara nilai USG pSWE dengan USG 2D-SWE, dan dilanjutkan dengan mencari formula regresi di antara kedua nilai tersebut.

Hasil: Pada lesi HCC terdapat korelasi positif kuat yang signifikan antara hasil USG pSWE dengan 2D-SWE pada perhitungan dengan kPa ($R = 0,882 / p < 0,01$) dan m/s ($R = 0,875 / p < 0,01$), didapatkan pula formula regresi nilai kPa pSWE = $2,99 + 0,75 \times \text{kPa 2D-SWE}$ dan nilai m/s pSWE = $0,31 + 0,82 \times \text{m/s 2D-SWE}$.

Kesimpulan: Pada lesi HCC, dapat dilakukan pemeriksaan nilai kekakuan jaringan menggunakan pSWE maupun 2D-SWE, baik menggunakan satuan kPa maupun m/s dengan hasil yang setara.

.....

Background: Liver cancer is the second most common cause of cancer death worldwide, and the incidence of 'primary' liver malignancies has increased significantly, particularly Hepatocellular Carcinoma / HCC. Characteristics of tissue in malignancy are usually denser / stiffer than normal tissue. Ultrasound with shear wave technology is a method of calculating tissue stiffness, currently there are two types; point shear wave elastography (pSWE) and two-dimensional shear wave elastography (2D-SWE). There have not been studies using these two types of shear waves in the same HCC lesions to see the congruence between the two values.

Objective: To determine the degree of correlation between the USG pSWE value and the 2D-SWE USG value on HCC.

Methods: A total of 17 study subjects with a diagnosis of HCC were subjected to pSWE ultrasound examination and followed by 2D-SWE (on the same day or a maximum of seven days thereafter) on HCC

lesions to determine the value of tissue stiffness (in kPa and m/s units). After that, a correlation analysis was carried out between the USG pSWE and USG 2D-SWE values, and continued by looking for the regression formula between the two values.

Results: In HCC lesions, there was a significant positive correlation between pSWE ultrasound results and 2D-SWE in the calculation with kPa ($R = 0.882 / p < 0.01$) and m / s ($R = 0.875 / p < 0.01$), also obtained the regression formula for the kPa pSWE value = $2.99 + 0.75 \times \text{kPa 2D-SWE}$ and the m/s pSWE value = $0.31 + 0.82 \times \text{m / s 2D-SWE}$.

Conclusion: In HCC lesions, tissue stiffness values can be examined using pSWE and 2D-SWE, using either kPa or m / s units with equivalent results.