

Pengembangan dan Validasi Metode Analisis Ivermectin dalam Volumetric Absorptive Microsampling menggunakan Kromatografi Cair Kinerja Ultra Tinggi Tandem Spektrometri Massa = Development and Validation of Ivermectin Quantification Method in Volumetric Absorptive Microsampling by Liquid Chromatography Tandem Mass Spectrometry

Salsabila, author

Deskripsi Lengkap: <https://lib.ui.ac.id/detail?id=20520038&lokasi=lokal>

Abstrak

Ivermectin merupakan obat antihelministik berspektrum luas yang digunakan untuk mengendalikan penyakit onchocerciasis dari parasit nematoda. Sebagai obat cacing, ivermectin didisain untuk memiliki kadar yang tinggi di saluran cerna, sehingga yang masuk sistemik tergolong rendah. Oleh karena kadar ivermectin yang sangat kecil, dibutuhkan metode yang sensitif dan selektif untuk analisis ivermectin dalam darah. Beberapa metode telah dikembangkan menggunakan plasma dan Dried Blood Spot, namun masih terdapat kekurangan karena terdapat efek hematokrit. Oleh karena itu, penelitian ini bertujuan untuk mengembangkan metode analisis ivermectin yang tervalidasi dengan baku dalam doramectin menggunakan Kromatografi Cair Kinerja Ultra Tinggi – tandem Spektrometri Massa (KCKUT-SM/SM). Deteksi menggunakan spektrometri massa triple quadrupole dengan mode electrospray ionization (ESI) positif. Matriks biologis yang digunakan yaitu darah utuh dalam Volumetric Absorptive Microsampling (VAMS), kemudian dipreparasi menggunakan metode pengendapan protein dengan campuran acetonitril dan metanol (1:1). VAMS dipilih karena tidak terdapat efek hematokrit, dibutuhkan volume sampel yang sedikit dan akurat, serta pengambilannya lebih efisien. Kondisi analisis optimum diperoleh menggunakan kolom Acquity® UPLC BEH C18 (2,1 x 100 mm; 1,7 m); fase gerak ammonium format 5 mM pH 3 dan acetonitril (10:90); laju alir 0,2 mL/menit; mode elusi isokratik selama 5 menit dengan suhu kolom 50C; deteksi menggunakan Multiple Reaction Monitoring (MRM) pada nilai m/z 892,41 > 569,5 untuk ivermectin dan 916,41 > 331,35 untuk doramectin. Metode analisis telah tervalidasi seluruhnya mengikuti pedoman dari US Food and Drug Administration tahun 2018. Nilai LLOQ 1 ng/mL dengan rentang konsentrasi linear pada 1 - 150 ng/mL.....Ivermectin is a broad-spectrum anthelmintic used to control onchocerciasis from nematode parasites. As an anthelmintic, ivermectin is designed to have high levels in the gastrointestinal tract, so that the systemic intake is relatively low. Due to the very small concentration of ivermectin, a sensitive and selective method is needed for the analysis of ivermectin in blood. Several methods have been developed using plasma and Dried Blood Spots, but there are still shortcomings due to hematocrit effects. Therefore, this study was conducted to develop a validated ivermectin analysis method with doramectin as the internal standard in using Ultra High Performance Liquid Chromatography-Tandem Mass Spectrometry. The analysis was performed using a triple quadrupole mass spectrometry with positive electrospray ionization (ESI) mode. The biological matrix used whole blood in Volumetric Absorptive Microsampling (VAMS) and extracted using the protein precipitation method with mixture of acetonitrile and methanol (1:1). VAMS has some advantages such as not affected by hematocrit, requires small and fixed volume of sample, also more efficient sampling process. The optimum conditions were obtained using the Acquity® UPLC BEH C18 column (2,1 x 100 mm; 1,7 m); flow rate was 0,2 mL/min; mobile phase was ammonium formate 5 mM pH

3 and acetonitrile (10:90); isocratic elusion for 5 minutes; Multiple Reaction Monitoring (MRM) detection with m/z values were 892.41 > 569.5 for ivermectin and 916,41 > 331,35 for doramectin. The method has been fully validated following the guidelines from US FDA (2018). LLOQ value was 1 ng/mL with a linear concentration range of 1 – 150 ng/mL.