

# Pengembangan pelacak DNA non-radioaktif pada tes dot blot dalam rangka pemantauan program pengendalian filariasis di Indonesia = Developing anon-radioactive brugia malayi DNA probe in a dot blot assay for monitoring the progress of the filariasis control program in Indonesia

Taniawati Supali, author

Deskripsi Lengkap: <https://lib.ui.ac.id/detail?id=91477&lokasi=lokal>

---

## Abstrak

Filariasis adalah penyakit infeksi yang disebabkan oleh cacing filaria pada pembuluh dan kelenjar limfe, dan ditularkan melalui gigitan nyamuk. Gejala klinik akut berupa demam beruikang, peradangan saluran atau kelenjar limfe, oedema dan gejala kronis berupa elefantiasis. Penyakit ini menyerang kelompok masyarakat yang aktif bekerja di daerah pedesaan sehingga dapat menurunkan produktivitas ekonomi suatu komunitas.

Di Indonesia lebih dari 20 juta penduduk tinggal di daerah endemis filariasis dan kira-kira 3-4 juta dari jumlah tersebut terinfeksi filariasis (Partono & Bintari, 1989). Dan ke-3 spesies cacing filaria yang menginfeksi manusia, *Brugia malayi* mempunyai penyebaran yang paling luas di Indonesia.

Program pengendalian filariasis telah dilakukan pemerintah sejak tahun 1970, melalui pemberian DEC secara massal pada penduduk yang tinggal di daerah endemis. Ada beberapa kendala dalam memantau keberhasilan program tersebut, yaitu: (1) Keengganan penduduk diambil darah malam berulang-ulang (2) Mahalnya biaya operasional pengambilan darah malam, dan (3) Pemeriksaan entomologis konvensional melalui pembedahan nyamuk langsung di bawah mikroskop di lapangan tidak dapat membedakan spesies larva parasit, terutama di daerah B. malayi terdapat bersamaan dengan parasit filaria hewan B. pahangi.

Dengan menggunakan bioteknologi telah dikembangkan pelacak DNA yang ditandai molekul radioaktif untuk parasit B. malayi (Piessens dkk., 1987), tetapi pelacak DNA radioaktif tersebut mahal, waktu paruh pendek, perlu latihan khusus untuk pemakaiannya, perlu pembuangan khusus dan berbahaya bagi pemakainya.

Akhir-akhir ini telah dikembangkan pelacak DNA yang ditandai molekul nonradioaktif, tetapi pelacak DNA non-radioaktif ini kurang sensitif dibandingkan dengan pelacak DNA radioaktif. Oleh karena itu tujuan penelitian ini adalah mengembangkan pelacak DNA B. malayi non-radioaktif yang spesifik dan sensitif pada tes dot blot dan dapat digunakan sebagai alat pemantau secara entomologis pada program pengendalian filariasis B. malayi di daerah-daerah endemis di Indonesia. Ada 7 persyaratan yang harus dipenuhi sebelum pelacak DNA digunakan sebagai alat pemantau, yaitu: (1) Harus dapat dihasilkan dalam jumlah yang tidak terbatas dengan waktu yang relatif singkat di Indonesia, (2) Harus stabil dalam kurun waktu lama, (3) Harus spesifik dan sensitif untuk parasit B. malayi, (4) Harus dapat bereaksi dengan B. malayi di Indonesia, (5) Harus sensitif untuk mendeteksi I larva infeksi B. malayi dalam nyamuk, (6) Tes dot blot harus dapat diulang, dan (7) Tes dot blot harus mudah dilakukan.

Sikuen DNA berulang *B. malayi* dari beberapa daerah di Indonesia telah dilakukan sebelum menentukan sikuen pelacak DNA *B. malayi* barn. Tiga strain *B. malayi* dari Indonesia, yaitu *B. malayi* strain Kendari (zoofilik subperiodik), *B. malayi* strain Bengkulu (zoofilik subperiodik), dan *B. malayi* strain Buton (antrofilik) telah dianalisis hasil sikuen DNA berulangnya dengan komputer untuk mendapatkan suatu konsensus sikuen DNA berulang *B. malayi* Indonesia.

<hr>

Filariasis is a disease resulting from an infection with a nematode parasite in lymph vessels and lymph nodes, and is transmitted by mosquito bites. The acute clinical manifestations of lymphatic filariasis are characterized by fever, lymphadenitis, retrograde lymphangitis, oedema and the chronic clinical manifestation is characterized by elephantiasis. The disease is predominantly affecting the young and the active working people in rural and slum areas, therefore it will decrease the economic productivity of the community significantly.

More than 20 million people live in endemic areas and approximately 3-4 million are currently infected. From the three species of parasites infecting man, *B. malayi* is the major and is widely distributed throughout the Indonesian islands.

Filariasis control program has been launched by the government since 1970, however, it met the following constraints in monitoring the progress of control program, (1) Poor participation in night blood collection from the people, (2) High costs of surveillance, and (3) Inappropriate technology in conventional entomological assessment to distinguish the infective larvae in vector mosquitoes.

In the last few years, new techniques for entomological assessment were explored using biotechnology. A radioactive *B. malayi* DNA probe was developed (Piessens et al., 1987) The radioactive labelled DNA probes are not suitable for field use because they are expensive, they have a short shelf life, and special training for handling the probes is imperative. Besides, laboratories arrangements for safe disposal are necessary.

Recently, non-radioactive labelled DNA probes have been developed but these probes were less sensitive compared to the radioactive labelled probes. Therefore, the objective of this experiment is to develop a new non-radioactive *B. malayi* DNA probe, which has more advantages than the conventional radiolabelled DNA probe, is specific and sensitive in a dot blot assay as a tool in entomological assessment for monitoring the progress of the filariasis control program in *B. malayi* infected areas of Indonesia. There are 7 requirements to be fulfilled before the probe can be widely used, such as:

1. The probe should be produced in a sufficient quantity in a relatively short period in Indonesia.
2. The probe should be stable.
3. The probe should be specific and sensitive for *B. malayi* parasite.
4. The probe should be able to hybridize with the Indonesian *B. malayi* strains.
5. The probe should be sensitive enough to detect 1 L3 in mosquitoes.
6. The dot blot assay should be reproducible.
7. The dot blot assay should be simple.

A new 25-mer (25 nucleotides) *B. malayi* DNA probe was designed by-comparing the consensus sequences of *B. malayi* to *B. pahangi*. In order to produce the probe in Indonesia, it needs to be cloned in the plasmid DNA (plasmid bluescript). However, it was found that the probe is too small to be an effective probe in a vector DNA.