

Praktik Kerja di Industri PT CKD Otto Pharmaceuticals Periode 31 Oktober - 28 Desember 2022, Studi Pembersihan untuk Metode Pencucian Alat-alat Gelas di Laboratorium Quality Control PT CKD Otto Pharmaceuticals = Work Practice in the PT CKD Otto Pharmaceuticals Industry 31 October - 28 December 2022 Period, Cleaning Study for Laboratory Glassware Washing Method at the Quality Control Laboratory of PT CKD Otto Pharmaceuticals

Astri Maulinda Sari, author

Deskripsi Lengkap: <https://lib.ui.ac.id/detail?id=9999920529615&lokasi=lokal>

Abstrak

Praktik laboratorium yang baik menuntut penggunaan peralatan gelas yang bersih. Hal ini dikarenakan hasil pengujian sangat dipengaruhi oleh lingkungan kerja yang bersih dan memadai. Studi pembersihan alat-alat gelas di laboratorium quality control PT CKD Otto Pharmaceuticals dilakukan untuk mengevaluasi bahwa pencucian secara manual terhadap alat-alat gelas telah optimal membersihkan alat-alat gelas dan tidak meninggalkan residu. Evaluasi dilakukan dengan melihat parameter pengujian pH, total organic carbon, dan conductivity dari air yang digunakan untuk membilas alat gelas yang telah dilakukan pencucian secara manual. Studi pembersihan metode pencucian alat-alat gelas di laboratorium memenuhi semua kriteria spesifikasi. Prosedur pencucian yang diterapkan pada semua alat-alat gelas di laboratorium quality control di PT CKD Otto Pharmaceuticals sudah efektif untuk membersihkan semua alat-alat gelas yang terbebas dari residu bahan aktif dan excipien.

..... Good laboratory practice requires the use of clean glassware. This is because the test results are strongly influenced by a clean and adequate work environment. A study on cleaning glassware in the quality control laboratory of PT CKD Otto Pharmaceuticals was conducted to evaluate that manually washing glassware has optimally cleaned glassware and left no residue. The evaluation was carried out by looking at the testing parameters for pH, total organic carbon, and the conductivity of the water used to rinse glassware that had been washed manually. The study of cleaning methods for washing glassware in the laboratory met all specification criteria. The washing procedure applied to all glassware in the quality control laboratory at PT CKD Otto Pharmaceuticals has been effective in cleaning all glassware free from residues of active ingredients and excipients.